




UNIVERSITY OF
SOUTH CAROLINA

ADDENDUM NO. 1

TO: ALL VENDORS
FROM: Reed Bickers 
SUBJECT: Sealed Bid # USC-IFB-1292-RB
OPENING DATE AND TIME: REMAIN UNCHANGED
DATE: June 20, 2008

This Addendum No. 1 modifies the Bid invitation only in the manner and to the extent as stated herein.

Item One: **SPECIFICATIONS:** Page 13, **K. Remove reference to "vendor supplied gas cabinets".** Gas cabinets for the toxic gases exist currently. **Add: A closed box with exhaust that houses the MFC's and their related valves & plumbing is acceptable.**

Item Two: **SPECIFICATIONS:** Page 13, **J. Add: The mechanical pump & chiller will be located close to the main chassis, in the same lab and not behind a wall in an equipment chase. Due to the restricted space the main chassis is to be compact. The controls and PC screen should fit in the 22" face.**

Item Three: **SPECIFICATIONS:** Page 13, **H. Add: A single EMO for the entire system on the main chassis is acceptable.**

BIDDER SHALL ACKNOWLEDGE RECEIPT OF ADDENDUM NO. 1 IN THE SPACE PROVIDED BELOW AND RETURN IT WITH THEIR SEALED BID RESPONSE. FAILURE TO DO SO MAY SUBJECT BID TO REJECTION.

Authorized Signature

Firm

Date